



Certificate of Conformity

1-2/F., South Block, Building A2, No.3, Keyan Road, Science City,
Guangzhou High-Tech Industrial Development Zone, Guangdong,
China

Te l: 86-20-32209330
www.i-testlab.com

No. ITL-200513002

The following products have been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with the EMC directive.

Applicant : **Foshan YiFeng Electric Industrial Co., Ltd.**

Address : No. 898, Gao Ming Avenue East, He Cheng Blvd. GaoMing District,
Foshan City, GuangDong Province,China.

Product name : **VACUUM TESTER**

Model No. : **V-BOX**

Technical data : 100-240V~, 50/60Hz, 40W

Test Report No. : 200513002

Test Standards	
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61000-3-2:2014	Electromagnetic compatibility (EMC) -- Part 3-2: Limits - Limits for harmonic current emissions (equipment input current \leq 16 A per phase)
EN 61000-3-3:2013	Electromagnetic compatibility (EMC) -- Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current \leq 16 A per phase and not subject to conditional connection



The test report was carried out from the submitted type samples of a product in compliance with the specification of the respective standards. The certificate holder has the right to fix the CE-mark for EMC directive on the product complying with the inspection samples